


Heat Disinfectant for Hemodialysis Unit Cito NovaClean 50%





Description

Cito NovaClean 50% solution is a high level disinfectant based on citric acid as a main ingredient. Cito NovaClean 50% is a chemo-thermal disinfectant and decalcification of hemodialysis machines in on process and effects on a variety of microorganisms (bacteria, fungi, viruses and spores). The synergistic effect of its components makes Cito NovaClean 50% a potent disinfectant solution.

Advantages

- Simultaneous disinfection and descaling
- Prevents the formation of biofilm and high efficiency in its removal
- Excellent descaling (carbonate, calcium, magnesium)
- Eliminates organic blood residues (fat and protein)
- Colorless and odorless
- Excellent stability
- High compatibility with dialysis machine (without causing corrosion and damage to the machine)
- This solution is composed of edible organic acids, so it does not cause any allergies and toxicity for the patient (safe)
- This solution has a high stability and decomposes into non-toxic substances in nature (biodegradable and environmentally friendly)
- Distinctive sterilization power with heating mode
- Efficacy Result
- Disinfection effect

Cito NovaClean 50% has superior bactericidal, fungicidal, sporcidal and antiviral effect in 80 -85 °C.

Mechanism of antimicrobial action

Destruction of the phospholipid layers in the cell membrane; disturbance of the intracellular pH balance; activity strongly increased by an increase in temperature.

Test Organisms	Efficacy/time contact/concentration
Sporicidal: Bacillus subtilis (ATCC 6633)	100%, 10 min, 2%
Bactericidal: Staphylococcus aureus (ATCC 6538)	
Esterichia coli (ATCC 10536)	100%, 10 min, 2%
Pseudomonas aeruginosa (ATCC 15442)	
Fungicidal: Candida albicans (ATCC 10231)	99.99%, 10 min, 2%

Decalcification effect


Excellent removal of inorganic deposits

• Mechanism of decalcifying action

Formation of Ca/Mg water soluble salt

• Cleaning effect

Cito NovaClean 50% can remove and prevent biofilm and organic matter





Corrosion effect

Ingredients of Cito NovaClean 50% do not damage to machine component and pipelines.

Compositions

Citric acid 50%, other organic acid to increase effectiveness 2%.

Instruction for use

Cito NovaClean 50% can be used all of hemodialysis machine with heat function. To use Cito NovaClean 50%, follow the instruction of manufacturer of the hemodialysis machines for heat disinfection with citric acid based disinfectant, or follow the recommends of hemodialysis machine maintenance engineers.

Warnings and Precautions

- Inspect the container before use.
- The solution inside the canister should be transparent and without any suspended particles, and its packaging should be healthy, otherwise it should not be used.
- In case of direct contact with the Cito NovaClean 50% rinse the contacted place immediately with plenty of warm water.
- In case of accidental ingestion of Cito NovaClean 50%, see a doctor.
- Wearing gloves and face shield is recommended during the use of Cito NovaClean 50%.
- In closed containers, store at 1 to 30 ° C away from direct sunlight.
- Keep Cito NovaClean 50% standing during storage and transport.
- Avoid throwing canister during transport.

Packaging

10 L canister

Shelf –life

18 months from manufactured date



Name: Nova pharmed chemia

Factory: 4nd Bustan, 4nd Golestan, West Hafez Blvd, Eshtehard, Industrial Zone, IRAN

Office: Fifth floor, No. 44, Saadat Abad Ave., 1998836841, Tehran, IRAN

Tel: +98(21)220 74001 Fax: +98(21)220 73103